



Evaluation test lane (double sampling, consent)

	Amphia	UMCU	Aruba	Aruba	Erasmus	RIVM	RIVM	RIVM
	completed	completed	completed	Not started	Completed	Completed-analysis	Ongoing	ongoing
Assay	BD veritor	Abbot Panbio	Abbott Panbio	Roche (SD biosensor Q test)	Roche (SD biosensor Q test)	Roche (SD biosensor Q test)	Quidel, sofia rapid Ag test	SD biosensor F test
Population size	350	1257	134	(200)	984	397	~670	~750
Prevalence (PCR tested)	4.8%	9.8%	32.8%	...	19.1%	13.9
sensitivity	94.1% ***	73.2% **	81.8%		86.6%	89.1%		
specificity	100%	100%	100%		99.5%	100%		

5.1.2e

et al NTVG 2020

Gremmels et al manuscript in prep



Working group



National Institute for Public Health and the Environment
Ministry of Health, Welfare and Sport




UMC Utrecht



Asociacion di Specialista
HOSPITAL DI ARUBA

- 5.1.2e
- 5.1.2e
- 5.1.2e
- 5.1.2e

ErasmusMC



AmphiA

- 5.1.2e
- 5.1.2e
- 5.1.2e