

**Van:** [redacted] <[redacted]@lchulpmiddelen.nl>  
**Verzonden:** vrijdag 4 september 2020 12:39  
**Aan:** [redacted] <[redacted]@minvws.nl>; [redacted] Pels Rijcken  
 <[redacted]@pelsrijcken.nl>; [redacted] <[redacted]@lchulpmiddelen.nl>; [redacted]  
 <[redacted]@lchulpmiddelen.nl>  
**Onderwerp:** FW: testresult of third lot MJH needed

Ha allen,

Zojuist ontving ik het derder testrapport vanuit China, het resultaat is een PASS.

De resultaten van dit lot 1276 zijn behoorlijk afwijkend van de resultaten die het RIVM rapporteerde, de 30 samples laten bij CNAS een constante hoge score zien op de penetratie-test.

Er wordt door Beconnected de opmerking bijgeplaatst dat dit om een andere fabrikant gaat, dit maken ze op aan de hand van de kwaliteits-briefjes die bijgesloten zijn bij de maskers. Inderdaad is het ticket bij lot 1276 anders dan de lots 1267 en 1277, zoals het rode stempel en de opmaak van de linker kolom

Afijn dat lijkt me iets dat MJH ons beter kan vertellen. Laten we (maandag) even overleggen hoe we hier verder mee om gaan.

Met vriendelijke groet,

[redacted]

**Van:** [redacted] <[redacted]@beconnected.com.cn>  
**Verzonden:** vrijdag 4 september 2020 12:17  
**Aan:** [redacted] <[redacted]@lchulpmiddelen.nl>  
**Onderwerp:** 答复: testresult of third lot MJH needed

Dear [redacted]

See attached report.  
 The difference is manufacturer from 1,2 batch an 3<sup>rd</sup> batch.  
 See quality ticket in side of bag attached, which mentioned on the report as well.

Kind regards,

[redacted]



[www.beconnected.world](http://www.beconnected.world)  
 Phone: + 86 (0) 21 62515599 x 8011  
 Microsoft Teams: [redacted]@beconnected.com.cn,  
 Skype: [redacted]  
 Address: [redacted]

[redacted]

地址：艾镁瑞(上海)贸易有限公司 上海市普陀区长寿路1118号A幢24 B/C/E