




## Who is Pasaca Capital?



- US Owned Private Equity firm offering advisory services and fund managing billions of dollars in assets from Institutional and Ultra High Net Worth investors across Asia and U.S.
- Investment Strategy is focused on identifying under-valued assets to exploit that discount via value gaps that exist between different equity markets
- Focus industries are ***Entertainment, Finance, Healthcare, Industrial-Manufacturing, Infrastructure, and TMT***
- Founding Partner is Charles Huang, Ph.D. formerly top-ranked China Equity Market Analyst for Credit Lyonnais Securities Asia (CLSA). Led 20+ IPOs and secondary offerings during his career at CSLA and BNP Paribas. Earned his Ph.D. and MBA from University of Strathclyde, Scotland



# Innova Medical Group

- Stem – Cell
- Stroke
- Heart Disease
- Arterial Health
- Cancer
- HPV / Viral Testing

# Innova Biotime Medical Antigen Rapid Test



## ANTIGEN – RAPID DIAGNOSTIC

1ml Sputum  
Nasal/Throat Swab  
Buccal Scrape

Point of Care –  
USFDA Applied EUA August 3<sup>rd</sup>  
Acknowledgement Letter August 19<sup>th</sup>  
EUA Expected Mid-Sept.

Current Production 1 million per day; will reach 2 million per day by Sept. 14  
Xiamen, China / Los Angeles, California

Relative Sensitivity:	72/75	96.00% (88.75%~99.17%)	
Relative Specificity:	220/220	100.00% (98.34%~100.00%)	
Accuracy:	292/295	98.98% (97.06%~99.79%)	
Total Results	75	220	295

\* 95% Confidence Interval

# Innova Biotime Medical Antigen Test


**BIOTIME** Xiamen Biotime Biotechnology Co., Ltd  
Add: 3F-4F, No. 188, Pingcheng South Road, Haicang Street, Haicang District, Xiamen, Fujian, 361026, P. R. China  
Tel: 86-592-6883156 Fax: 86-592-6882362 Web: www.biotime.cn

## Brochure



# QDX™ Health ID Smartphone App

## A Robust, Highly Secure Health ID Application




- The entire system is US HIPAA compliant, uses HL7 data compliance and is supported by blockchain security
- Tracking and reporting within the app managed privately and securely
- Objective: keep contagious travelers, customers and staff from contaminating the greater population in the airport, manufacturing facility and community
- Adaptive protocols and enterprise dashboard

# QDX™ HealthID


*A Key or Passport to Healthy Havens*

**QDX™ HealthID** is a secure, trusted, always-on service that provides end-to-end visibility to support testing and immunization for infectious diseases on a global scale, as well as associated back-to-work, play and life for individuals that are verifiably clear of a disease.

## Simple-Authentic-Certified



# What Does QDX™ HealthID Do?



## What QDX™ HealthID Does *NOT* Do?



No Storage of Personal Data



No Storage of Test Results



No Storage Employment Status Decisions



No Decision-Making Authority on Health Status

## COVIDLytics™ Explained

*A combination of analytical modeling and encrypted "health passport" app leveraging Innova's integrated tests*

- Analytical modeling which captures global infections rates daily on a detailed level - country by country, state by state, etc., and uses that data to perform computational and highly accurate modeling for any given test environment
- Medical data, symptoms, and actual testing information using antibody or antigen testing, or in combination
- Ideal for large campuses, manufacturing, airports, airlines, cruise lines
- Objective: analytics used to quantify risk in a specific environment on an individual basis and building up to a risk profile which will not overwhelm the COVID health response in any particular area



# Healthy Havens for Work, Travel, Play & Living



# Appendix

## SARS-CoV-2 Antigen Rapid Qualitative Test

### Instructions for Use

#### For prescription only

#### For in vitro diagnostic use only

Please read these instructions completely before beginning testing of specimens.

#### INTENDED USE

The SARS-CoV-2 Antigen Rapid Qualitative Test is a colloidal gold immunochromatography intended for the qualitative detection of nucleocapsid antigens from SARS-CoV-2 in human nasal swabs or throat swabs from individuals who are suspected of COVID-19 by their healthcare provider, within the first five days of the onset of symptoms.

Results are for the identification of SARS-CoV-2 nucleocapsid antigen. Antigen is generally detectable in upper respiratory samples or lower respiratory samples during the acute phase of infection. Positive results indicate the presence of viral antigens, but clinical correlation with patient history and other diagnostic information is necessary to determine infection status. Positive results do not rule out bacterial infection or co-infection with other viruses. The agent detected may not be the definite cause of disease.

Negative results do not rule out SARS-CoV-2 infection and should not be used as the sole basis for treatment or patient management decisions, including infection control decisions. Negative results should be considered in the context of a patient's recent exposures, history and the presence of clinical signs and symptoms consistent with COVID-19, and confirmed with a molecular assay, if necessary for patient management.

The SARS-CoV-2 Antigen Rapid Qualitative Test is intended for use by trained clinical laboratory personnel specifically instructed and trained in the techniques of in vitro diagnostic procedures, and proper infection control procedures and individuals similarly trained in point of care settings.

#### SUMMARY

SARS-CoV-2 belongs to the broad family of viruses known as coronaviruses. It is a positive-sense single-stranded RNA (+ssRNA) virus. Other coronaviruses are capable of causing illnesses ranging from the common cold to more severe diseases such as Middle East respiratory syndrome (MERS). It is the seventh known coronavirus to infect people, after 229E, NL63, OC43, HKU1, MERS-CoV, and the original SARS-CoV. Protein modeling experiments on the spike (S) protein of the virus suggest that it has sufficient affinity to the angiotensin converting enzyme 2 (ACE2) receptors of human cells to use them as a mechanism of cell entry. Studies have shown that SARS-CoV-2 has a higher affinity to human ACE2 than the original SARS virus strain.

SARS-CoV-2 infections cause COVID-19 disease. People who have confirmed COVID-19 have a range of symptoms, from people with little to no symptoms to people being severely sick and dying. Symptoms can include fever, tiredness, and dry cough. Some patients may have aches and pains, nasal congestion, runny nose, sore throat or diarrhea. These symptoms are usually mild and begin gradually. Some people become infected but don't develop any symptoms and don't feel unwell. Most people (about 80%) recover from the disease without needing special treatment. Around 1 out of every 6 people who gets COVID-19 becomes seriously ill and develops difficulty breathing. Older people, and those with underlying medical problems like high blood pressure, heart problems or diabetes, are more likely to develop serious illness. About 2% of people with the disease have died. People with fever, cough and difficulty breathing should seek medical attention.

Human-to-human transmission of the virus has been confirmed and occurs primarily via respiratory droplets from coughs and sneezes within a range of about 6 feet (1.8m). Viral RNA has also been found in stool specimens from infected patients. It is possible that the virus can be infectious even during the incubation period, but this has not been proven, and the WHO stated on 1 February 2020 that "transmission from asymptomatic cases is

likely not a major driver of transmission" at this time.

The median incubation time is estimated to be approximately 5 days with symptoms estimated to be present within 12 days of infection. The symptoms of COVID-19 are similar to other viral respiratory diseases and include fever, cough, shortness of breath.

#### PRINCIPLES OF THE PROCEDURE

This reagent is based on colloidal gold immunochromatography assay.

During the test, specimen extracts are applied to the test cartridges. If there were SARS-CoV-2 antigen in the extract, the antigen will bind to the SARS-CoV-2 monoclonal antibody. During lateral flow, the complex will move along the nitrocellulose membrane toward the end of the absorbent paper. When passing the test line (line T, coated with another SARS-CoV-2 monoclonal antibody) the complex is captured by SARS-CoV-2 antibody on test line resulting in coloring on line T, when passing the line C, colloidal gold-labeled goat anti-rabbit IgG is captured by control line (line C, coated with rabbit IgG) resulting in coloring on line C.

#### REAGENTS

The following components are included in the SARS-CoV-2 Antigen Rapid Qualitative Test for rapid detection of SARS-CoV-2 kit:

#### Materials Provided:

25-Test Kit:

1. SARS-CoV-2 Antigen Test Cartridge (25) Monoclonal anti-SARS antibodies
2. Extraction Tubes (25)
3. Extraction solution: 2 bottles/kit (enough for 25 test)
4. Instructions for use 1 copy/kit
5. QC Card (located on kit box)

#### Optional Materials:

1. Throat Swabs (25)
2. Nasal Swabs (25)

#### Materials Required but not provided:

1. Timer
2. Tube rack for specimens
3. Any necessary personal protective equipment
4. External control set (including 1 negative controls and 1 positive controls).

#### WARNINGS AND PRECAUTIONS

1. For in vitro diagnostic use.
2. This test has been authorized only for the detection of proteins from SARS-CoV-2, not for any other viruses or pathogens.
3. Do not use this kit beyond the expiration date printed on the outside carton.
4. Do not use the kit to evaluate patient specimens if either the positive control swab or negative control swab fail to give expected results.

## Innova Medical Group Inc.

5. Test results are meant to be visually determined.

6. To avoid erroneous results, specimens must be processed as indicated in the assay procedure section.

7. Do not reuse any kit components.

8. When collecting a nasal swab sample, use the nasal swab supplied in the kit. Use of alternative swabs may result in false negative results.

9. Proper specimen collection, storage and transport are critical to the performance of this test.

10. Specific training or guidance is recommended if operators are not experienced with specimen collection and handling procedures. Wear protective clothing such as laboratory coats, disposable gloves, and eye protection when specimens are collected and evaluated. Pathogenic microorganisms, including hepatitis viruses and Human Immunodeficiency Virus, may be present in clinical specimens. Standard precautions and institutional guidelines should always be followed in handling, storing, and disposing of all specimens and all items contaminated with blood or other body fluids.

11. The SARS-CoV-2 external positive control have been prepared from recombinant viral proteins and do not contain infectious material.

12. Dispose of used test kits as biohazardous waste in accordance with federal, state and local requirements.

13. For additional information on hazard symbols, safety, handling and disposal of the components within this kit, please refer to the Safety Data Sheet (SDS) located at [bd.com](http://bd.com).

14. Wear suitable protective clothing, gloves, and eye/face protection when handling the contents of this kit.

#### STORAGE CONDITIONS & PERIOD OF VALIDITY

1. Store extraction solution at 2-30°C, the shelf life is 24 months tentatively.
2. Store the test cartridge at 2-30°C, the shelf life is 24 months tentatively.
3. Test Cartridge should be used right after opening the pouch.

Reagents and devices must be at room temperature (15-30 °C) when used for testing.

#### SPECIMEN COLLECTION AND HANDLING

##### Specimen Collection and Preparation

##### Throat Swab Specimen Collection:

Let the patient's head tilt slightly, mouth open, and make "ah" sounds, exposing the pharyngeal tonsils on both sides. Hold the swab and wipe the pharyngeal tonsils on both sides of the patient with moderate force back and forth for at least 3 times.



##### Nasal Swab Specimen Collection:

1. Insert the swab into one nostril of the patient. The swab tip should be inserted up to 2.5


INNOVA

## Innova Medical Group Inc.

cm (1 inch) from the edge of the nostril.

2. Roll the swab 5 times along the mucosa inside the nostril to ensure that both mucus and cells are collected.

3. Using the same swab, repeat this process for the other nostril to ensure that an adequate sample is collected from both nasal cavities. Withdraw the swab from the nasal cavity.

**Specimen Transport and Storage:**

Samples should be tested as soon as possible after collection. Based on data generated with influenza virus, throat swabs are stable for up to 24-hours at room temperature or 2° to 8°C.

**TEST METHODS**

The test should be operated at room temperature (15-30°C).

1. Place the extraction tube with opening facing up. Press the extraction solution bottle to drip 6 drops of extract solution into the extractor tube without touching the edge of the tube.


2. The extraction of specimen: Put the swab had collected specimen into the extraction tube, hold and press the swab head against the wall of tube with force while rotating the swab for about 10 seconds to release the antigen into the extraction solution from the swab head.

3. Removing swab: Squeeze the swab head while removing the swab in order to remove as much liquid as possible from the swab. Dispose of swabs according to biohazard waste disposal regulations.

4. Install the nozzle cap onto the extraction tube.

5. Loading: drip 2 drops of extraction solution into the sample well of the test cartridge, and start the timer.


6. Read the results at 20-30 minutes. Strong positive results can be reported at 20 minutes, however, negative results must be reported after 30 minutes. If positive signal appears after 30 minutes, it should not be reported as positive.



Drip extract solution

Extraction

Removing swab



Install the nozzle

Loading

Reading

**INTERPRETATION OF TEST RESULTS**


Line C must be colored to have a valid test result.

**Valid results:**

**Negative result:** There is coloration on line C only showing as following picture, suggesting that there is no SARS-CoV-2 antigen in the specimen.

**Positive result:** There are coloration on both line C and line T showing as follow pictures, suggesting that there is SARS-CoV-2 antigen in the specimen.

**Invalid result:** There is no coloration on line C, as shown in the following pictures. The test is invalid or an error in operation occurred. Repeat the assay with a new cartridge.



Negative

Positive

Invalid

**REPORTING OF RESULTS****Positive Test:**

Positive for the presence of SARS-CoV-2 antigen. Positive results indicate the presence of viral antigens, but clinical correlation with patient history and other diagnostic information is necessary to determine infection status. Positive results do not rule out bacterial infection or co-infection with other viruses. The agent detected may not be the definite cause of disease. Laboratories within the United States and its territories are required to report all positive results to the appropriate public health authorities.

**Negative Test:**

Negative results are presumptive. Negative test results do not preclude infection and should not be used as the sole basis for treatment or other patient management decisions, including infection control decisions, particularly in the presence of clinical signs and symptoms consistent with COVID-19, or in those who have been in contact with the virus. It is recommended that these results be confirmed by a molecular testing method, if necessary, for patient management. Control.

**Invalid:**

Do not report results. Repeat the test.

**QUALITY CONTROL**

The SARS-CoV-2 Antigen Control Set (catalog number: 1339) is available to purchase separately from Xiamen Biotime Biotechnology Co., Ltd as external controls. The control set can be ordered through website ([www.biotime.cn](http://www.biotime.cn)), telephone (+86-592-4883156) and email ([baotai@biotime.cn](mailto:baotai@biotime.cn)). One negative and one positive control are included in the control set. Returning expected test results for each control in the control set indicates appropriate performance of SARS-CoV-2 Antigen Rapid Qualitative Test. If any control of the control set fail to provide the expected result, reasons that have led to failure including the test kit, the operator, the environment, the test procedure and any other causes which may affect the test result should be analyzed and corrective action taken. Clinical specimens can be run in the Biotime SARS-CoV-2 Antigen Rapid Qualitative Test. If all the control set results observed are the expected results. Please refer to the Instructions For Use of Biotime SARS-CoV-2 Antigen Control Set for expected test results as well as other information. It is recommended that the controls are tested when:

- A new operator uses the kit;
  - A new lot of test kits is used;
  - A new shipment of kits is used;
  - The temperature used during storage of the kit falls outside of the recommended conditions;
  - The temperature of the test area falls outside of 15-30°C;
  - To verify a higher than expected frequency of positive or negative results;
  - To investigate the cause of repeated invalid results; or
  - A new test environment is used (e.g., natural light vs. artificial light).
- I. As required by external quality control procedures and in accordance with local, state and federal regulations or accreditation requirements.

**NOTE:** Prepare kit control swabs and test using the same procedure as used for patient specimens. Failure of the external/procedural controls will generate an invalid test result.

If the kit controls do not perform as expected, do not report patient results. Contact Xiamen Biotime Biotechnology Co., Ltd Technical Services at (+86-592-6883577) and email ([baotai@biotime.cn](mailto:baotai@biotime.cn)).

**LIMITATIONS OF THE PROCEDURE**

- Clinical performance was evaluated with frozen samples, and test performance may be different with fresh samples.
- Users should test specimens as quickly as possible after specimen collection.
- Positive test results do not rule out co-infections with other pathogens.
- Results from SARS-CoV-2 Antigen Rapid Qualitative Test should be correlated with the clinical history, epidemiological data, and other data available to the clinician evaluating the patient.
- A false-negative test result may occur if the level of viral antigen in a sample is below the detection limit of the test or if the sample was collected or transported improperly; therefore, a negative test result does not eliminate the possibility of SARS-CoV-2 infection.
- The amount of antigen in a sample may decrease as the duration of illness increases. Specimens collected after day 5 of illness are more likely to be negative compared to a RT-PCR assay.

INNOVA

7.Failure to follow the test procedure may adversely affect test performance and/or invalidate the test result.

8.The contents of this kit are to be used for the qualitative detection of SARS-CoV-2 antigens from throat or nasal swab specimens only.

9.The kits for rapid detection of SARS-CoV-2 can detect both viable and non-viable SARS-CoV-2 material. The SARS-CoV-2 Antigen Rapid Qualitative Test for rapid detection of SARS-CoV-2 performance depends on antigen load and may not correlate with other diagnostic methods performed on the same specimen.

10.Negative test results are not intended to rule in other non-SARS-CoV-2 viral or bacterial infections.

11.Positive and negative predictive values are highly dependent on prevalence rates. Positive test results are more likely to represent false positive results during periods of little/no SARS-CoV-2 activity when disease prevalence is low. False negative test results are more likely when prevalence of disease caused by SARS-CoV-2 is high.

12.This device has been evaluated for use with human specimen material only.

13.Monoclonal antibodies may fail to detect, or detect with less sensitivity, SARS-CoV-2 viruses that have undergone minor amino acid changes in the target epitope region.

14.The performance of this test has not been evaluated for use in patients without signs and symptoms of respiratory infection and performance may differ in asymptomatic individuals.

15.Sensitivity of the test after the first five days of the onset of symptoms has been demonstrated to decrease as compared to a RT-PCR SARS-CoV-2 assay.

16.The kit was validated with the assorted swabs. Use of alternative swabs may result in false negative results.

17.Specimen stability recommendations are based upon stability data from influenza testing and performance may be different with SARS-CoV-2. Users should test specimens as quickly as possible after specimen collection, and within one hour after specimen collection.

18.The validity of SARS-CoV-2 Antigen Rapid Qualitative Test has not been proven for identification/confirmation of tissue culture isolates and should not be used in this capacity.

#### CLINICAL PERFORMANCE

The performance of the Biotme SARS-CoV-2 Antigen Rapid Qualitative Test for rapid detection of SARS-CoV-2 was established with 295 direct nasal swab or throat swab prospectively collected and enrolled from individual symptomatic patients (within 5 days of onset) who were suspected of COVID-19. As with all antigen tests, performance may decrease as days since symptom onset increases. For each type, four kinds of samples from the same person were tested by Company's Kit. We selected 25 positive and 25 negative sample. P1-P25 of samples are from infected people, and N1-N25 are from uninfected people. P21-P25 are weekly positive.

Method	PCR Test			Total Results
	Results	positive	Negative	
Biotme Results	positive	72	0	72
	Negative	3	220	223

Relative Sensitivity:	72/75	96.00% (88.75%~99.17%)
Relative Specificity:	220/220	100.00% (98.34%~100.00%)
Accuracy:	292/295	98.98% (97.06%~99.79%)

Total Results	75	220	295
---------------	----	-----	-----

\* 95% Confidence Interval

#### ANALYTICAL PERFORMANCE

##### LIMIT OF DETECTION (ANALYTICAL SENSITIVITY)

##### LOD of human sputum matrix

The LOD for the SARS-CoV-2 Antigen Rapid Qualitative Test for rapid detection of SARS-CoV-2 was established using limiting dilutions of heat-inactivated SARS-CoV-2 antigen (bei Resources NR-52286). The material was supplied frozen at a concentration of TCID50 of  $3.40 \times 10^5$  per mL.

In this study, designed to estimate the LOD of the assay when using a direct throat swab, the starting material was spiked into a volume of pooled human sputum obtained from healthy donors and confirmed negative for SARS-CoV-2. An initial range finding study was performed testing devices in triplicate using a 10-fold dilution series of 3 replicates per concentration. At each dilution, 50  $\mu$ L samples were added to swabs and then tested in the assay using the procedure appropriate for patient throat swab specimens. A concentration was chosen between the last dilution to give 3 positive results and the first to give 3 negative results. Using this concentration, the LOD was further refined with a 2-fold dilution series of 3 replicates per concentration. The last dilution demonstrating 100% positivity was then tested in an additional 20 replicates tested in the same way.

Starting Material Concentration	Estimated LOD	No. Positive/Total	% Positive
$3.40 \times 10^5$ TCID50/mL	$4.25 \times 10^7$ TCID50/mL	19/20	95%

##### LOD of human nasal swab specimen matrix

The LOD for the SARS-CoV-2 Antigen Rapid Qualitative Test for rapid detection of SARS-CoV-2 was established using limiting dilutions of heat-inactivated SARS-CoV-2 (bei Resources NR-52286). The material was supplied frozen at a concentration of TCID50 of  $3.40 \times 10^5$  per mL.

In this study, designed to estimate the LOD of the assay when using a direct throat swab, the starting material was spiked into a volume of pooled human nasal swab specimen obtained from healthy donors and confirmed negative for SARS-CoV-2. An initial range finding study was performed testing devices in triplicate using a 10-fold dilution series of 3 replicates per concentration. At each dilution, 50  $\mu$ L samples were added to swabs and then tested in the assay using the procedure appropriate for patient nasal swab specimens. A concentration was chosen between the last dilution to give 3 positive results and the first to give 3 negative results. Using this concentration, the LOD was further refined with a 2-fold dilution series of 3 replicates per concentration. The last dilution demonstrating 100% positivity was then tested in an additional 20 replicates tested in the same way.

#### Innova Medical Group Inc.

Starting Material Concentration	Estimated LOD	No. Positive/Total	% Positive
$3.40 \times 10^5$ TCID50/mL	$3.40 \times 10^5$ TCID50/mL	19/20	95%

##### CROSS REACTIVITY (ANALYTICAL SPECIFICITY)

##### Human sputum matrix

Cross-reactivity of the SARS-CoV-2 Antigen Rapid Qualitative Test for rapid detection of SARS-CoV-2 was evaluated by testing a panel of high-prevalence respiratory pathogens that could potentially cross-react with the SARS-CoV-2 Antigen Rapid Qualitative Test for rapid detection of SARS-CoV-2. Each organism and virus spiked into negative throat specimen was wet-tested in triplicate. The final concentration of each organism is documented in the following table.

S.N.	Potential Cross-Reactant	Concentration Tested	Cross-Reactivity (Yes/No)
1	Human coronavirus 229E	$2.0 \times 10^5$ TCID50/mL	NO
2	Human coronavirus OC43	$2.0 \times 10^5$ TCID50/mL	NO
3	Human coronavirus NL63	$2.0 \times 10^5$ TCID50/mL	NO
4	SARS-coronavirus	$2.0 \times 10^5$ TCID50/mL	NO
5	MERS-coronavirus	$2.0 \times 10^5$ TCID50/mL	NO
6	Human coronavirus HKU1	$2.0 \times 10^5$ TCID50/mL	NO
7	Adenovirus C1	$2.0 \times 10^5$ TCID50/mL	NO
8	Adenovirus 71	$2.0 \times 10^5$ TCID50/mL	NO
9	Human Metapneumovirus (hMPV)	$2.0 \times 10^5$ TCID50/mL	NO
10	Parainfluenza virus 1-4	$2.0 \times 10^5$ TCID50/mL	NO
11	Influenza A	$2.0 \times 10^5$ TCID50/mL	NO
12	Influenza B	$2.0 \times 10^5$ TCID50/mL	NO
13	Enterovirus	$2.0 \times 10^5$ TCID50/mL	NO
14	Respiratory syncytial virus	$2.0 \times 10^5$ TCID50/mL	NO
15	Rhinovirus	$2.0 \times 10^5$ TCID50/mL	NO
16	Haemophilus influenzae	$2.0 \times 10^6$ TCID50/mL	NO
17	Streptococcus pneumoniae	$2.0 \times 10^6$ TCID50/mL	NO

INNOVA

## Innova Medical Group Inc.

18	Streptococcus pyogenes	$2.0 \times 10^6$ TCID50/mL	NO
19	Candida albicans	$2.0 \times 10^6$ TCID50/mL	NO
S.N.	Potential Cross-Reactant	Concentration Tested	Cross-Reactivity (Yes/No)
20	Pooled human nasal wash – representative of normal respiratory microbial flora	$2.0 \times 10^6$ TCID50/mL	NO
21	Bordetella pertussis	$2.0 \times 10^6$ TCID50/mL	NO
22	Mycoplasma pneumoniae	$2.0 \times 10^6$ TCID50/mL	NO
23	Chlamydia pneumoniae	$2.0 \times 10^6$ TCID50/mL	NO
24	Legionella pneumophila	$2.0 \times 10^6$ TCID50/mL	NO
25	Mycobacterium tuberculosis	$2.0 \times 10^6$ TCID50/mL	NO
26	Pneumocystis jirovecii (PJP)	$2.0 \times 10^6$ TCID50/mL	NO

Note: 1 TCID50/mL ≈ 0.7 CFU/ml

Based on the data generated by this study, the substances tested SARS-CoV-2 Antigen Rapid Qualitative Test do not cross-react.

## Human nasal swab specimen matrix

Cross-reactivity of the SARS-CoV-2 Antigen Rapid Qualitative Test for rapid detection of SARS-CoV-2 was evaluated by testing a panel of high prevalence respiratory pathogens that could potentially cross-react with the SARS-CoV-2 Antigen Rapid Qualitative Test for rapid detection of SARS-CoV-2. Each organism and virus spiked into negative nasal specimen was wet-tested in triplicate. The final concentration of each organism is documented in the following table.

S.N.	Potential Cross-Reactant	Concentration Tested	Cross-Reactivity (Yes/No)
1	Human coronavirus 229E	$2.0 \times 10^5$ TCID50/mL	NO
2	Human coronavirus OC43	$2.0 \times 10^5$ TCID50/mL	NO
3	Human coronavirus NL63	$2.0 \times 10^5$ TCID50/mL	NO
4	SARS-coronavirus	$2.0 \times 10^5$ TCID50/mL	NO
5	MERS-coronavirus	$2.0 \times 10^5$ TCID50/mL	NO
6	Human coronavirus HKU1	$2.0 \times 10^5$ TCID50/mL	NO

7	Adenovirus C1	$2.0 \times 10^5$ TCID50/mL	NO
8	Adenovirus 71	$2.0 \times 10^5$ TCID50/mL	NO
9	Human Metapneumovirus (hMPV)	$2.0 \times 10^5$ TCID50/mL	NO
10	Parainfluenza virus 1-4	$2.0 \times 10^5$ TCID50/mL	NO
11	Influenza A	$2.0 \times 10^5$ TCID50/mL	NO
12	Influenza B	$2.0 \times 10^5$ TCID50/mL	NO
13	Enterovirus	$2.0 \times 10^5$ TCID50/mL	NO
14	Respiratory syncytial virus	$2.0 \times 10^5$ TCID50/mL	NO
15	Rhinovirus	$2.0 \times 10^5$ TCID50/mL	NO
16	Haemophilus influenzae	$2.0 \times 10^6$ TCID50/mL	NO
17	Streptococcus pneumoniae	$2.0 \times 10^6$ TCID50/mL	NO
18	Streptococcus pyogenes	$2.0 \times 10^6$ TCID50/mL	NO
19	Candida albicans	$2.0 \times 10^6$ TCID50/mL	NO
20	Pooled human nasal wash – representative of normal respiratory microbial flora	$2.0 \times 10^6$ TCID50/mL	NO
21	Bordetella pertussis	$2.0 \times 10^6$ TCID50/mL	NO
22	Mycoplasma pneumoniae	$2.0 \times 10^6$ TCID50/mL	NO
23	Chlamydia pneumoniae	$2.0 \times 10^6$ TCID50/mL	NO
24	Legionella pneumophila	$2.0 \times 10^6$ TCID50/mL	NO
25	Mycobacterium tuberculosis	$2.0 \times 10^6$ TCID50/mL	NO
26	Pneumocystis jirovecii (PJP)	$2.0 \times 10^6$ TCID50/mL	NO

Note: 1 TCID50/mL ≈ 0.7 CFU/ml

Based on the data generated by this study, the substances tested SARS-CoV-2 Antigen Rapid Qualitative Test do not cross-react.

## MICROBIAL INTERFERENCE STUDIES

## Human sputum matrix

The microbial interference studies for the SARS-CoV-2 Antigen Rapid Qualitative Test for rapid detection of SARS-CoV-2 was established using limiting dilutions of heat-inactivated SARS-CoV-2 (Bei Resources NR-52286).

The material was supplied frozen at a concentration of TCID50 of  $3.40 \times 10^5$  per mL. The starting material was spiked into a volume of pooled human sputum (the most challenging respiratory matrix) obtained from healthy donors and confirmed negative for SARS-CoV-2. Based on the LOD studies, a low (3x LoD) SARS-CoV-2 concentration of  $1.275 \times 10^5$  TCID50/mL was chosen. The specimen was confirmed positive for SARS-CoV-2 with family line on Line T. Furthermore, the above-mentioned specimen was divided into 30. Finally, the microorganism indicated below was respectively spiked into the divided specimen to obtain microbial interference specimens that SARS-CoV-2 is present in a specimen with one microorganism.

Each microbial interference specimen was tested individually. At each test, 50 µL samples were added to swab. The results show that the specimen was confirmed positive for SARS-CoV-2 with family Line on Line T. Based on the study, no appreciable interference was observed for the following substances at the spiked levels indicated below in sputum matrix.

S.N.	Potential Cross-Reactant	Concentration Tested	Cross-Reactivity (Yes/No)
1	Human coronavirus 229E	$2.0 \times 10^5$ TCID50/mL	NO
2	Human coronavirus OC43	$2.0 \times 10^5$ TCID50/mL	NO
3	Human coronavirus NL63	$2.0 \times 10^5$ TCID50/mL	NO
4	SARS-coronavirus	$2.0 \times 10^5$ TCID50/mL	NO
5	MERS-coronavirus	$2.0 \times 10^5$ TCID50/mL	NO
6	Human coronavirus HKU1	$2.0 \times 10^5$ TCID50/mL	NO
7	Adenovirus C1	$2.0 \times 10^5$ TCID50/mL	NO
8	Adenovirus 71	$2.0 \times 10^5$ TCID50/mL	NO
9	Human Metapneumovirus (hMPV)	$2.0 \times 10^5$ TCID50/mL	NO
10	Parainfluenza virus 1-4	$2.0 \times 10^5$ TCID50/mL	NO
11	Influenza A	$2.0 \times 10^5$ TCID50/mL	NO
12	Influenza B	$2.0 \times 10^5$ TCID50/mL	NO
13	Enterovirus	$2.0 \times 10^5$ TCID50/mL	NO
14	Respiratory syncytial virus	$2.0 \times 10^5$ TCID50/mL	NO
15	Rhinovirus	$2.0 \times 10^5$ TCID50/mL	NO

INNOVA

## Innova Medical Group Inc.

16	Haemophilus influenzae	2.0 x 10 <sup>6</sup> TCID50/mL	NO
17	Streptococcus pneumoniae	2.0 x 10 <sup>6</sup> TCID50/mL	NO
18	Streptococcus pyogenes	2.0 x 10 <sup>6</sup> TCID50/mL	NO
19	Candida albicans	2.0 x 10 <sup>6</sup> TCID50/mL	NO
20	Pooled human nasal wash – representative of normal respiratory microbial flora	2.0 x 10 <sup>6</sup> TCID50/mL	NO
21	Dodeletella pertussis	2.0 x 10 <sup>6</sup> TCID50/mL	NO
22	Mycoplasma pneumoniae	2.0 x 10 <sup>6</sup> TCID50/mL	NO
23	Chlamydia pneumoniae	2.0 x 10 <sup>6</sup> TCID50/mL	NO
24	Legionella pneumophila	2.0 x 10 <sup>6</sup> TCID50/mL	NO
25	Mycobacterium tuberculosis	2.0 x 10 <sup>6</sup> TCID50/mL	NO
26	Pneumocystis jirovecii (PJP)	2.0 x 10 <sup>6</sup> TCID50/mL	NO

**Human nasal swab specimen matrix**

The microbial interference studies for the SARS-CoV-2 Antigen Rapid Qualitative Test for rapid detection of SARS-CoV-2 was established using limiting dilutions of heat-inactivated SARS-CoV-2 (see Resources NR-52286).

The material was supplied frozen at a concentration of TCID50 of 3.40 x 10<sup>5</sup> per mL, the starting material was spiked into a volume of pooled nasal swab specimen (the most challenging respiratory matrix) obtained from healthy donors and confirmed negative for SARS-CoV-2. Based on the LOD studies, a low (3x LoD) SARS-CoV-2 concentration of 1.02 x 10<sup>3</sup> TCID50/mL was chosen. The specimen was confirmed positive for SARS-CoV-2 with faint line on Line T. Furthermore, the above-mentioned specimen was divided into 30. Finally, the microorganism indicated below was respectively spiked into the divided specimen to obtain microbial interference specimens that SARS-CoV-2 is present in a specimen with one microorganism.

Each microbial interference specimen was tested individually. At each test, 50 µL samples were added to swab. The results show that the specimen was confirmed positive for SARS-CoV-2 with faint line on Line T. Based on the study, no appreciable interference was observed for the following substances at the spiked levels indicated below in nasal swab specimen matrix.

S.N.	Potential Cross-Reactant	Concentration Tested	Cross-Reactivity (Yes/No)
1	Human coronavirus 229E	2.0 x 10 <sup>5</sup> TCID50/mL	NO
2	Human coronavirus OC-43	2.0 x 10 <sup>5</sup> TCID50/mL	NO

3	Human coronavirus NL63	2.0 x 10 <sup>6</sup> TCID50/mL	NO
4	SARS-coronavirus	2.0 x 10 <sup>5</sup> TCID50/mL	NO
5	MERS-coronavirus	2.0 x 10 <sup>5</sup> TCID50/mL	NO
6	Human coronavirus HKU1	2.0 x 10 <sup>5</sup> TCID50/mL	NO
7	Adenovirus C1	2.0 x 10 <sup>5</sup> TCID50/mL	NO
8	Adenovirus 71	2.0 x 10 <sup>5</sup> TCID50/mL	NO
9	Human Metapneumovirus (hMPV)	2.0 x 10 <sup>5</sup> TCID50/mL	NO
10	Parainfluenza virus 1-4	2.0 x 10 <sup>5</sup> TCID50/mL	NO
11	Influenza A	2.0 x 10 <sup>5</sup> TCID50/mL	NO
12	Influenza B	2.0 x 10 <sup>5</sup> TCID50/mL	NO
13	Enterovirus	2.0 x 10 <sup>5</sup> TCID50/mL	NO
14	Respiratory syncytial virus	2.0 x 10 <sup>5</sup> TCID50/mL	NO
15	Rhinovirus	2.0 x 10 <sup>5</sup> TCID50/mL	NO
16	Haemophilus influenzae	2.0 x 10 <sup>6</sup> TCID50/mL	NO
17	Streptococcus pneumoniae	2.0 x 10 <sup>6</sup> TCID50/mL	NO
18	Streptococcus pyogenes	2.0 x 10 <sup>6</sup> TCID50/mL	NO
19	Candida albicans	2.0 x 10 <sup>6</sup> TCID50/mL	NO
20	Pooled human nasal wash – representative of normal respiratory microbial flora	2.0 x 10 <sup>6</sup> TCID50/mL	NO
21	Bordetella pertussis	2.0 x 10 <sup>6</sup> TCID50/mL	NO
22	Mycoplasma pneumoniae	2.0 x 10 <sup>6</sup> TCID50/mL	NO
23	Chlamydia pneumoniae	2.0 x 10 <sup>6</sup> TCID50/mL	NO
24	Legionella pneumophila	2.0 x 10 <sup>6</sup> TCID50/mL	NO

25	Mycobacterium tuberculosis	2.0 x 10 <sup>6</sup> TCID50/mL	NO
26	Pneumocystis jirovecii (PJP)	2.0 x 10 <sup>6</sup> TCID50/mL	NO

**Endogenous Interference Substances Studies:****Human sputum matrix**

A study was performed demonstrate that eighteen (18) potentially interfering substances that may be found in the lower respiratory tract do not cross-react or interfere with the detection of SARS-CoV-2 in the SARS-CoV-2 Antigen Rapid Qualitative Test.

S.N	Interfering Substance	Concentration	Cross-Reacting Results	Interference Results**
1	Whole Blood	4%	Negative	Positive
2	Mucin	0.50%	Negative	Positive
3	Ricola (Menthol)	1.5 mg/mL	Negative	Positive
4	Secrets (Dyclonin/Menthol)	1.5 mg/mL	Negative	Positive
5	Chloraseptic (Menthol/Benzocaine)	1.5 mg/mL	Negative	Positive
6	Naso GEL (NeilMed)	5% v/v	Negative	Positive
7	CVS Nasal Drops (Phenylephrine)	15% v/v	Negative	Positive
8	Afrin (Oxymetazoline)	15% v/v	Negative	Positive
9	CVS Nasal Spray (Tremolyn)	15% v/v	Negative	Positive
10	Nasal Gel (Oxymetazoline)	10% v/v	Negative	Positive
11	Zicam	5% v/v	Negative	Positive
12	Homeopathic (Alkalol)	1:10 dilution	Negative	Positive
13	Fisherman's Friend	1.5 mg/mL	Negative	Positive
14	ore Throat Phenol Spray	15% v/v	Negative	Positive
15	Tobramycin	4µg/mL	Negative	Positive



16	Mupirocin	10 mg/mL	Negative	Positive
17	Fluticasone Propionate	5% v/v	Negative	Positive
18	Tamiflu (Oseltamivir Phosphate)	5mg/mL	Negative	Positive

Based on the data generated by this study, the substances tested SARS-CoV-2 Antigen Rapid Qualitative Test do not cross-react or interfere.

#### Human nasal swab specimen matrix

A study was performed demonstrate that eighteen (18) potentially interfering substances that may be found in the upper respiratory tract do not cross-react or interfere with the detection of SARS-CoV-2 in the SARS-CoV-2 Antigen Rapid Qualitative Test.

S.N	Interfering Substance	Concentration	Cross- Reactive Results	Interference Results**
1	Whole Blood	4%	Negative	Positive
2	Mucin	0.50%	Negative	Positive
3	Ricola (Menthol)	1.5 mg/mL	Negative	Positive
4	Sucrets (Dyclonin/Menthol)	1.5 mg/mL	Negative	Positive
5	Chlorasptic (Menthol/Perizocaine)	1.5 mg/mL	Negative	Positive
6	Naso GEL (NasalMed)	5% v/v	Negative	Positive
7	CVS Nasal Drops (Phenylephrine)	15% v/v	Negative	Positive
8	Afrin (Oxymetazoline)	15% v/v	Negative	Positive
9	CVS Nasal Spray (Cromolyn)	15% v/v	Negative	Positive
10	Nasal Gel (Oxymetazoline)	10% v/v	Negative	Positive
11	Zicam	5% v/v	Negative	Positive
12	Homeopathic (Alkalol)	1:10 dilution	Negative	Positive
13	Fisherman's Friend	1.5 mg/mL	Negative	Positive
14	ore Throat Phenol Spray	15% v/v	Negative	Positive
15	Tobramycin	4.11 g/mL	Negative	Positive


















16	Mupirocin	10 mg/mL	Negative	Positive
17	Fluticasone Propionate	5% v/v	Negative	Positive
18	Tamiflu (Oseltamivir Phosphate)	5mg/mL	Negative	Positive

Based on the data generated by this study, the substances tested SARS-CoV-2 Antigen Rapid Qualitative Test do not cross-react or interfere.

#### HIGH DOSE HOOK EFFECT

As part of the LoD study the highest concentration of heat-inactivated SARS-CoV-2 stock available (TCID<sub>50</sub> of 3.40 x10<sup>7</sup> per mL) was tested. There was no Hook effect detected.

#### INDEX OF SYMBOLS

Symbol	Description	Symbol	Description
	In vitro diagnostic medical device		Do not re-use
	Expiry date		Consult instructions for use
	Warning, please refer to the instruction		Manufacturer
	Store at 2-30°C		Lot number
	Keep away from sunlight		Keep dry
	European authorized representative		Don't use the product when the package is damaged
	Date of manufacture		Biological risks
	For Prescription Only		CE mark
	Sterilized using ethylene oxide		

## Innova Medical Group Inc.

### IN VITRO DIAGNOSTIC MEDICAL DEVICE TECHNICAL ASSISTANCE

For technical assistance, call Biotime Technical Services at +86 592 688 3577, email [51126@biotime.cn](mailto:51126@biotime.cn), or visit Biotime website at <http://www.biotime.cn>.

### GENERAL INFORMATION

Manufactured by:

**Xiamen Biotime Biotechnology Co., Ltd**

Address: 3F/4F, No. 188, Pingcheng South Road, Haicang District, Xiamen, Fujian, 361026, P.R. China

Tel: +86-592-6883577

Fax: +86-592-6882362

[www.biotime.cn](http://www.biotime.cn)

Manufactured for: **Innova Medical Group Inc.**



**Innova Medical Group Inc.**

Address: 718 S. Primrose Ave, Menlo Park, CA 91016, USA

Tel: 626-239 0025

Fax: 626-239 0038

[www.innovamedgroup.com](http://www.innovamedgroup.com)

Version: A/02  
Issuing date: 2020-07-01

## **SARS-CoV-2-Antigen Rapid Qualitative Test Kit Clinical Report**

**Project Name:**SARS-CoV-2-Antigen Rapid Qualitative Test Kit clinical test

**Testing Time:**2020.01~2020.7

**Statistical Unit (Signature and seal):** Xiamen Biotime Biotechnology Co.,Ltd.

5.1.2e

**Address:** No.188, Pingchengnan Road, Haicang Street, Haicang District,  
Xiamen City

SARS-CoV-2-Antigen Rapid Qualitative Test Kit Clinical Report

---

## 目 录

1. Experimental design.....	2
1. 1 Overall design and scheme description of the test .....	2
1. 2 Test Methods .....	2
1. 3 Test sample requirements .....	2
1. 4 Sample Test.....	3
1. 4. 1 Test flow chart.....	3
2. Clinical trial results and analysis .....	3
2. 1 Basic data statistics .....	3
2. 2 Homology comparison analysis.....	3
2. 2. 1 Homology comparison test results .....	3
3. Conclusion.....	5

## 1. Experimental design

### 1.1 Overall design and scheme description of the test

Blind data analysis was used in this clinical study, Use Xiamen Biotime Biotechnology Co., Ltd. Manufacture SARS-CoV-2-Antigen Rapid Qualitative Test Kit (Colloidal Gold). The detection results of the samples were compared with PCR method for comparative study. After the end of the test, the blinding was uncovered, the cause of the inconsistencies was analyzed, and all the inconsistencies in the test should be fully analyzed in combination with the patient's epidemiological background, clinical symptoms, disease outcome and other information.

### 1.2 Test Methods

(1) The nasal swabs of 75 patients with novel Coronavirus nucleic acid positive and 220 subjects with novel Coronavirus nucleic acid negative were tested simultaneously using the product and PCR method to evaluate the sensitivity, specificity and accuracy of the product.

(2) Homologous pharyngeal swab samples were collected from the above samples: 25 positive samples, 20 weakly positive samples, and 25 negative samples were required to evaluate the consistency of nasal swabs and pharyngeal swabs.

### 1.3 Test sample requirements

#### 1.3.1 Sample type

The samples were nasal swabs and pharyngeal swabs.

#### 1.3.2 Inclusion criteria for clinical trial samples


(1) Confirmed cases of pneumonia infected by novel Coronavirus: patients with positive nucleic acid from novel Coronavirus were selected.

(2) Confirmed cases of pneumonia infected with pneumonia other than novel Coronavirus: patients with nucleic acid negative of Novel Coronavirus were selected.

## SARS-CoV-2-Antigen Rapid Qualitative Test Kit Clinical Report

## 1. 4 Sample Test

## 1. 4. 1 Test flow chart



## 2. Clinical trial results and analysis

## 2. 1 Basic data statistics

The sensitivity, specificity and accuracy were calculated and the 95% confidence intervals were calculated respectively.

Method		PCR Test		Total Results
	Results	positive	Negative	
Biotime KIT	positive	72	0	72
	Negative	3	220	223
Total Results		75	220	295

Relative Sensitivity:	72/75	96.00% (88.75%~99.17%)
Relative Specificity:	220/220	100.00% (98.34%~100.00%)
Accuracy:	292/295	98.98% (97.06%~99.79%)

## 2. 2 Homology comparison analysis

## 2. 2. 1 Homology comparison test results

Sample NO.	Nasal swab	Throat swab
P1	+++	+++
P2	+	+
P3	++	+

## SARS-CoV-2-Antigen Rapid Qualitative Test Kit Clinical Report

P4	+	+
P5	++	++
P6	+	+
P7	++	+
P8	+	+
P9	++	++
P10	+	+
P11	+++	++
P12	+	+
P13	++	++
P14	++	+
P15	+	+
P16	+++	+++
P17	+	+
P18	+++	+++
P19	++	++
P20	+++	++
P21	+	+
P22	+	+
P23	+	+
P24	+	+
P25	+	+
N1	-	-
N2	-	-
N3	-	-
N4	-	-
N5	-	-
N6	-	-
N7	-	-
N8	-	-
N9	-	-

## SARS-CoV-2-Antigen Rapid Qualitative Test Kit Clinical Report

N10	-	-
N11	-	-
N12	-	-
N13	-	-
N14	-	-
N15	-	-
N16	-	-
N17	-	-
N18	-	-
N19	-	-
N20	-	-
N21	-	-
N22	-	-
N23	-	-
N24	-	-
N25	-	-

Annotation: P1-P25 of samples are from infected people , and NI-N25 are from uninfected people. P21-P25 are weekly positive.

### 3. Conclusion

The consistency between PCR method and this product is good. There was no difference between nasal and pharyngeal swabs.

# CE Declaration of Conformity CE

## Manufacturer

Xiamen Biotime Biotechnology Co., Ltd.  
3F/4F, No.188, Pingcheng South Road, Haicang Street, Haicang District, Xiamen,  
Fujian 361026, P. R. China

## Declaration of Conformity

We, Xiamen Biotime Biotechnology Co., Ltd., hereby declare that the below mentioned medical device meets the provisions of Directive 98/79/EC which apply to them. The declaration of conformity is exclusively under the responsibility of Xiamen Biotime Biotechnology Co., Ltd.

### (A) Particulars of medical device

Product name: SARS-CoV-2 Antigen Rapid Qualitative Test  
Manufacturing Site: 3F/4F, No.188, Pingcheng South Road, Haicang Street, Haicang District, Xiamen, Fujian 361026, P. R. China  
Risk-based classification: Other  
Route to compliance: Annex III of Directive 98/79/EC

### (B) Quality Management System

QMS certificate: ISO 13485 certificate  
Certificate number: CN 17/42021  
Issuance date: 7 December 2017  
Expiry date: 6 December 2020  
Conformity Assessment Body: SGS United Kingdom Ltd

### (C) Authorized Representative:

Logcon East GmbH.  
1110 Vienna Austria, Europe

5.1.2e

(Authorized Signature)

Name 5.1.2e

Title 5.1.2e

Place: Xiamen

Date of Issue: July 11, 2020



## Acknowledgment Letter

8/19/2020

5.1.2e

Xiamen Biotime Biotechnology Co., Ltd  
3F/4F, No. 188, Pingcheng S. Road, Haicang District  
Xiamen, Fujian 361026  
CHINA

Dear 5.1.2e:

The Center for Devices and Radiological Health (CDRH) of the Food and Drug Administration (FDA) has received your submission. This submission has been assigned the unique document control number below. All future correspondence regarding this submission should be identified prominently with the number assigned and should be submitted to the Document Control Center at the above letterhead address. Failure to do so may result in processing delays. If you believe the information identified below is incorrect, please contact the Office of Product Evaluation and Quality (OPEQ) submission support at (301) 796-5640 or

5.1.2e [@fda.hhs.gov](mailto:5.1.2e@fda.hhs.gov).

Submission Number: EUA202568  
Received: 8/19/2020  
Applicant: Xiamen Biotime Biotechnology Co., Ltd  
Device: SARS-CoV-2 Antigen Rapid Qualitative Test

We will notify you when the review of this document has been completed or if any additional information is required. For information about CDRH review regulations and policies, please refer to <http://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/default.htm>.

Sincerely yours,

Center for Devices and Radiological Health