

To: [redacted] 5.1.2e [redacted] 5.1.2e @rivm.nl; [redacted] 5.1.2e [redacted] 5.1.2e @rivm.nl; [redacted] 5.1.2e @erasmusmc.nl; [redacted] 5.1.2e @erasmusmc.nl; [redacted] 5.1.2e @umcutrecht.nl; [redacted] 5.1.2e @umcutrecht.nl; [redacted] 5.1.2e @rivm.nl; [redacted] 5.1.2e @erasmusmc.nl; [redacted] 5.1.2e @erasmusmc.nl; [redacted] 5.1.2e @gmail.com; [redacted] 5.1.2e @gmail.com]
Cc: [redacted] 5.1.2e @roche.com; [redacted] 5.1.2e [redacted] 5.1.2e @roche.com]
From: [redacted] 5.1.2e
Sent: Mon 9/21/2020 3:33:50 PM
Subject: Roche Diagnostics SARS-CoV-2 Rapid Antigen test
Received: Mon 9/21/2020 3:34:59 PM
[Method Sheet SARS-CoV-2 Rapid Antigen Test.pdf](#)
[Infographics Rapid antigen test.pdf](#)
[QRG SARS-CoV-2 Rapid Antigen Test.pdf](#)
[Overleg verificatie SARS-CoV-2 Rapid antigen test final.docx](#)

Dear all,

First of all, thank you for your time and being available at very short notice for a consultation and the pleasant conversation. In the attachment you find conversation notes

In the meantime, I have some answers to questions you have asked.

1. About the inactivation of the virus, we have received the following information:

The manufacturer is currently conducting tests on the topic of virus inactivation.

We have seen interim results which have shown a virus inactivation in the buffer solution after from 2 minutes.

We will have to wait for the final report, I will sent it to you ASAP.

2. About the second question: Can the test be done in the lab where samples are collected in the test lane and transported to the lab within x hours? Or is the test onsite, meaning test should be done in less than 1-2 hours

In the Instruction for Use is the following statement:

8. The sample should be tested as soon as possible after collection.

9. Samples may be stored at room temperature for up to 1 hour or at 2 8 °C/ 36 46 °F for up to 4 hours prior to testing.

3. The Test procedure you will find in the Method Sheet and the infographic, attached to this email.

I hope to have been of sufficient service, if you have any further questions and / or comments don't hesitate to ask.

Kind Regard

[redacted] 5.1.2e

Met vriendelijke groet / kind regards,

[redacted] 5.1.2e

[redacted] 5.1.2e

[redacted] 5.1.2e

Roche Diagnostics Nederland BV
 Centralized Diagnostics
 Transistorstraat 41
 1322 CK Almere The Netherlands

office +31 [redacted] 5.1.2e
 mobile +31 [redacted] 5.1.2e

[redacted] 5.1.2e @roche.com

www.roche.nl

twitter: RocheDiagnosticsNL

Confidentiality Note: This message is intended only for the use of the named recipient(s) and may contain confidential and/or privileged information. If you are not the intended recipient, please contact the sender and delete this message. Any unauthorized use of the information contained in this message is prohibited.

