

Serial #: HFG4N08HY0

Alarm Status

OK 

Monitor Configuration

Programmer: admin  
Start Up Delay: 30 min Interval: 5 min

Monitor Read

On : 24-SEP-2020 2:00:27 PM By : TempTale® Ultra  
Note: All Times in GMT

Recorded Data

First Point: 22-SEP-2020 4:57:51 AM  
Stop Time: 23-SEP-2020 9:28:56 PM  
Number of Points: 487  
Trip Length: 1 day 16 hr 35 min

Summary Data

Low Extreme: 19.4 °C @ 23-SEP-2020 8:27:51 PM  
High Extreme: 20.6 °C @ 22-SEP-2020 4:57:51 AM  
Mean ± Std Deviation: 20.1 °C ± 0.3 °C  
Mean Kinetic Temperature: 20.2 °C  
Activation Energy: 83.144 kJ/mol

Alarm Summary

Alarm	Alarms Settings	Degree Minutes	Threshold (Type)	Total Time	Number of Events	Triggered	Status
Ideal range	>= 15.0 °C <= 25.0 °C	--	Unlimited	1 d 16 h 35 m	1	--	--
Alarm 1	Over 25.0 °C	0.0 °C min	0 d 0 h 0 m (S/E)	0 d 0 h 0 m	0	--	OK
Alarm 2	Under 15.0 °C	0.0 °C min	0 d 0 h 0 m (S/E)	0 d 0 h 0 m	0	--	OK

Alarm Data

Originator Notes:

