

To: 5.1.2e [5.1.2e @rivm.nl]
Cc: 5.1.2e [5.1.2e @sciensano.be]; 5.1.2e [5.1.2e @sciensano.be]
From: 5.1.2e
Sent: Wed 9/23/2020 12:30:05 PM
Subject: RE: oracol swabs en SARS-CovID
Received: Wed 9/23/2020 12:30:18 PM

Hi 5.1.2e

Thank you very much for your quick reply.
Very interesting findings and happy to discuss this further with 5.1.2e

For PCR testing, you perform two Oracol samples in order to obtain sufficient saliva (minimal 2ml).
We tested it by ourselves and indeed it is not easy to collect 1 ml with only 1 Oracol device.

Probably I will come back to you with some extra questions.

Kind regards
5.1.2e

5.1.2e 5.1.2e 5.1.2e
5.1.2e

Viral Diseases
National Reference Center for Measles, Mumps, Rubella, Viral Hepatitis and Human Papillomavirus
T + 32 5.1.2e


Rue Juliette Wytsmanstraat 14 • 1050 Brussels
Laboratory adress : Rue Engelandstraat 642 • 1180 Brussels
www.sciensano.be

 Please consider the environment before printing this e-mail

