

Reporting Form	
Laboratory:	Triskelion (Ducares BV)
Performer:	5.1.2e
Reporter:	5.1.2e
Date panels received: 08-10-2020 (2020.2) 13-10-2020 (2020.3, 2020.6)	
Date panels tested: 15-10-2020	
Date panel results reported: 15-10-2020	
In the question below use the official equipment name or the reagent/kit name as is being used by the manufacturer	
One-step real-time RT-PCR	
RNA/NA extraction	
Equipment: PurePrep Nucleic Acid Purification System (Molgen)	
Reagents/kit: PurePrep Pathogens (Molgen)	
One-step real-time RT-PCR:	
Kit for RT-PCR: Rida Gene SARS CoV 2 (PG6815)	
In-house method, nvt	
Primers/probes reference:nvt	
Enzyme(s):nvt	
Buffer:nvt	
PCR Equipment:ABI7500	
Other nucleic acid detection method	
Equipment:	
Reagents/kit:	
Volumes	
Classical one-step real-time RT-PCR or other NAAT that needs nucleic acid extraction	
Volume specimen in nucleic acid extraction: 200 µl	elution in volume: 100 µl
Volume RNA/total NA in RT-PCR/other NAAT reaction:5 µl	end volume RT-PCR/other NAAT reaction: 25 µl
Other nucleic acid detection method	
Volume specimen input:	

Results					
Please report Ct values or units if the work flow generate these					
Specificity panel SARS-CoV-2 (panel number from the box: 2020.6)					
Panel member number	Ct value/units Target gene 1: E gene SARS CoV 2	Ct value/units Target gene 2: nvt	Ct value/units Target gene 3: Internal Control	Conclusion (Pos/Neg/ Indeterminate/ Equivocal)	Comment
8. EQA6 CoV20-01	undetermined		31.2	NEG	
9. EQA6 CoV20-02	undetermined		31.2	NEG	
10. EQA6 CoV20-03	undetermined		31.4	NEG	
11. EQA6 CoV20-04	36.0		31.1	Indeterminate	
12. EQA6 CoV20-05	undetermined		31.6	NEG	
13. EQA6 CoV20-06	36.5		31.2	Indeterminate	
14. EQA6 CoV20-07	undetermined		31.2	NEG	
15. EQA6 CoV20-08	undetermined		31.1	NEG	
16. EQA6 CoV20-09	undetermined		30.9	NEG	
17. EQA6 CoV20-10	29.7		31.0	POS	
Sensitivity panel SARS-CoV-2 (panel number from the box: 2020.2)					
1. Sen. Series-01	undetermined		31.5	NEG	
2. Sen. Series-02	29.6		30.6	POS	
3. Sen. Series-03	undetermined		30.8	NEG	
4. Sen. Series-04	26.9		30.2	POS	
5. Sen. Series-05	undetermined		31.5	NEG	
6. Sen. Series-06	34.4		31.3	POS	
7. Sen. Series-07	37.3		31.1	Indeterminate	
Sensitivity panel SARS-CoV-2 (panel number from the box: 2020.3)					
18. Sen. Series3-01	30.5		31.0	POS	
19. Sen. Series3-02	undetermined		31.1	NEG	
20. Sen. Series3-03	41.3		31.4	NEG	
21. Sen. Series3-04	40.2		31.2	NEG	
22. Sen. Series3-05	34.1		31.0	POS	

23. Sen. Series3-06	undetermined		31.3	NEG	
24. Sen. Series3-07	26.7		30.2	POS	
Other comments:					
<ul style="list-style-type: none">• CT < 35 = POS• CT > 35 - <38 = Indeterminate• CT > 38 = NEG					
If you have used more than one work flow, provide the results of each work flow in a separate reporting form.					

Send report to: [@rivm.nl](mailto:5.1.2e@rivm.nl)

Many thanks for completing and submitting the results!